

**UNDER POSTING OF CERTIFICATE**



SARDAR VALLABHBHAI NATIONAL  
INSTITUTE OF TECHNOLOGY,  
SURAT (GUJARAT) 395007.  
APPLIED CHEMISTRY DEPARTMENT

Phone (O): 0261-2223371 to 74  
Office (ASHD): 0261-2201542  
Fax: 0261-2227334/2228394  
Grams: SVARCET

**ENQUIRY FOR SUPPLY OF ANALYTICAL WEIGHING BALANCE**

Ref. No: ACD/Annual Plant/Analytical Balance/1648/ ~~1707~~

Date: 05.03.2019


To


Sub: Enquiry for Supply of ANALYTICAL WEIGHING BALANCE  
Dear Sir/Madam

You are invited to quote your prices for supply of stores listed overleaf. The quotations may be sent in a sealed envelope with the superscription "Quotation with reference to the Enquiry No. ACD/Annual Plant/Analytical Balance/1648/ ~~1707~~ date 05.03.2019. Attention: Dr. B. Z. Dholakiya, Head, Department of Applied Chemistry, and addressed to The Director, S. V. National Institute of Technology, Surat – 395 007. **Your quotation should reach the under signed on or before 06.04.2019.** With quotations furnish the following information

1. The brand or make of an item should be specifically stated and wherever necessary. Complete set of specifications and dimensions should be given.
2. If asked, samples are to accompany the quotations.
3. Sales tax, General tax, Central Sales tax, Custom duty, Insurance charges, packing and forwarding charges, if not included under the price quoted, should be clearly specified.
4. The period of validity of the quotation must be specified. Offers subject to prior sale may please be avoided.
5. The delivery period is to be clearly mentioned in the quotation.
6. The mode of delivery of the stores may be mentioned. The delivery should be F.O.R Surat or the Institute. The institute is located in the Municipal Limits and exempted from paying of Octroi Duty.
7. All concessions available to an educational institution should be specified and also taken into account while quoting.
8. Payment will be made only after satisfactory of the supplied item.
9. The institute is registered with the Department of Scientific and Industrial Research (DSIR) for the purpose of availing customs duty exemption and central excise duty exemption and hence the certificate to this effect will be issued wherever it is necessary for demand.
10. Payment is normally made by cheque drawn on the S.V.R.C.E.T Branch Office of State Bank of India, Surat – 395 007 within a period of thirty days from the date of receipt of stores.
11. Your specifications and terms conditions should be on your company letterhead & signed by an authorized person.
12. Offered Quotation may be rejected if any ambiguity is found in the offered specifications, terms & conditions supplied by the party in specified tabular format.
13. The director reserves the right to accept stores which are not strictly in confirming with the specifications, but otherwise found suitable.

Yours faithfully

  
(Dr. K. Suresh Kumar)  
In-charge, Analytical Weighing Balance


  
Dr. B. Z. Dholakiya  
Head, ACD


**Head**  
Applied Chemistry Department  
S. V. National Institute of Technology  
SURAT-395007.

**SPECIFICATIONS OF ANALYTICAL WEIGHING BALANCE TO BE SUPPLIED**

Sr. No	Item	Specifications	Quantity	Quoted amount including all taxes & other charges (amount in rupees)
1	<b>Analytical Weighing Balance</b>	1. Complete with glass wind shield and one touch. 2. Function key having sliding door opening/closing system. 3. Overload Protection 4. Leveling feet and Level indicator 5. Weigh between 0.01 mg to 220g with a readability of 0.01 mg 6. Built-in calibration weight 7. Repeatability : 0.01 mg. 8. Linearity : 0.2 mg. 9. Response time : 2-3 sec. 10. AC Adapter 230 Volts. 11. Tare range full. 12. Platform dimension: 80 - 90 mm. 13. Protective Covers. 14. LCD display 15. Provision of indication when the sample weight reached the set point. 16. Operating temperature 5-50°C. 17. Provided with non-reactive pan, spatula, cleaner and cover for dust protection etc. 18. Users list regarding supply during last two years.	03	

**Note:** Parties must mention make & model no. of equipment offered by them with detail specification (on separate sheet).

  
(Dr. K. Suresh Kumar)  
In-charge, Analytical Weighing Balance

  
Dr. B. Z. Dholakiya  
Head, ACD

**Head**  
Applied Chemistry Department  
S. V. National Institute of Technology  
SURAT-395007.