



Offi : 0261- 2201532

Fax : 0261-2228394

Grams : SVNIT

**Sardar Vallabhbhai National Institute of Technology,
Surat-395 007.**

No.: SVNIT/DCE/AK/2021-22/ 5181

Date: 03/03/2022

SUB: - Enquiry for Supply of Slake Durability Apparatus (Refer Annexure-1)

Dear Sir,

You are requested to quote your prices for supply of stores listed overleaf. The quotations may be sent to the undersigned (Head, DCE) in a sealed envelope and subscribed as: **“Quotation with reference to Enquiry No. SVNIT/DCE/AK/2021-22/ 5181 dated 03/03/2022 on or before 11/03/2022.**

The quotations should be furnished with the following information.

- 1) The brand or make of each item should be specifically stated and wherever necessary, Complete set of specifications and dimensions should be given.
- 2) If asked, samples are accompanying the quotations
- 3) Sales tax, General tax, Central Sales tax, Custom duty, Delivery and Installation charges, if not included in the prices quoted, should be clearly specified.
- 4) The period of validity of the quotation should be at least 60 Days. Offers subject to prior sale may please be avoided.
- 5) The delivery period (maximum 30 days is allowed for this item from date of issue of supply order) is to be clearly mentioned in the quotation.
- 6) The mode of delivery of the stores may be mentioned. The delivery should be F.O.R. SVNIT, Surat.
- 7) All concessions available to an educational institution should be specified and also taken into account while quoting.
- 8) This Institute is located within the limits of S.M.C. & exempted from the paying of octroi duty on incoming goods from outside limits of S.M.C.
- 9) This Institute is registered with the Dept. of Scientific & Industrial Research (DSIR) for the purpose of availing custom duty exemption & central excise duty exemption, and hence the certificate to this effect will be issued wherever it is necessary on demand.
- 10) **Please mention the warranty period (Minimum one year warranty is essential)**
- 11) Payment is normally made by cheque drawn on the S.V.N.I.T. Branch Office of State Bank of India, Surat-395007 within a period of thirty days from the date of successful installation of the stores.
- 12) The Director reserves the right to accept stores, which are not strictly in confirming with the specifications but otherwise, found suitable.

Yours faithfully,

Kumar. A.
03/03/2022



**Assistant Professor
Department of Civil Engineering
S.V.N.I.T., Surat-395 007.**

Head
Department of Civil Engineering
SVNIT, Surat-395 007

**Professor & Head
Department of Civil Engineering
S.V.N.I.T., Surat-395 007. 1**

Annexure-1

Sr. No.	Item/Stores with Specifications	Probable quantity required
1	Slake Durability Apparatus <ol style="list-style-type: none">1. Mortar capable of rotating four drums at 20 rev/min.2. Two test drums of brass wire mesh with one end closed and having a removable cover at other end.3. Diameter 140 mm, unobstructed length 100 mm.4. Two perspex water troughs accommodating one test drum each.5. The troughs allow 40 mm unobstructed clearance below the test drum. A water level mark is provided in each water trough 20 mm below the axis of rotation of the drum.6. Suitable for rock testing.7. Suitable for operation on 220 v, 50 hz, single phase supply.	1

Note- The quotation not fulfilling strictly the above technical specification will be rejected.

Kumar. A.
03/03/2022

Assistant Professor
Department of Civil Engineering
S.V.N.I.T., Surat-395 007.

