



Offi : 0261- 2201842
Fax : 0261-2227334
Grams : SVNIT
Sardar Vallabhbhai National Institute of Technology,
Surat-395 007.

No.: SVNIT/DoCE/S Patel/2026-27/ 2136

Date: 30/07/2026
31

SUB: - Enquiry for Supply of Vibrating Wire Earth Pressure Cell


Dear Sir,

You are requested to quote your prices for supply of stores listed overleaf. The quotations may be sent to the undersigned (Head, DoCE) in a sealed envelope and subscribed as: "Quotation with reference to Enquiry No.SVNIT/DoCE/S. Patel/2026-27/..2136.... dated ..31/7/26...on or before 10/08/2026.

The quotations should be furnished with the following information.

- 1) The brand or make of each item should be specifically stated and wherever necessary, Complete set of specifications and dimensions should be given.
- 2) If asked, samples are accompany the quotations
- 3) Sales tax, General tax, Central Sales tax, Custom duty, Delivery and Installation charges, if not included in the prices quoted, should be clearly specified.
- 4) The period of validity of the quotation should be at least 60 Days. Offers subject to prior sale may please be avoided.
- 5) The delivery period (maximum 60 days is allowed for this item from date of issue of supply order) is to be clearly mentioned in the quotation.
- 6) The mode of delivery of the stores may be mentioned. The delivery should be F.O.R. SVNIT, Surat.
- 7) All concessions available to an educational institution should be specified and also taken into account while quoting.
- 8) Please mention the warranty period (Minimum one year warranty is essential)
- 9) Payment is normally made by cheque drawn on the S.V.N.I.T. Branch Office of State Bank of India, Surat-395007 within a period of thirty days from the date of successful installation of the stores.
- 10) The Director reserves the right to accept stores, which are not strictly in confirming with the specifications but otherwise, found suitable.

Yours faithfully,


30/07/2026
Dr. Satyajit Patel




Head
Department of Civil Engineering
SVNIT, Surat-395 007



(2)

Sr. No.	Item/Stores with Specifications	Probable quantity required	Whether sample of paper/Register/Item is to be sent or not
1	Vibrating Wire Earth Pressure Cell - Sensor element: Vibrating wire sensor - Range: 5kg/cm ² - Resolution: 0.025% FSR - Accuracy: $\pm 0.5\%$ FSR - Nonlinearity: $\pm 0.5\%$ FSR - Over range capacity: 150% FSR - Built-in temperature device: Thermistor (3k Ω) - Material: Stainless steel, high grade epoxy resin potting - Dimension: $\Phi 100\text{mm} \times 420(\text{L})\text{mm}$ (Customized) Signal cable - $\Phi 6.4\text{mm}$, $0.37\text{mm}^2 \times 4\text{C}$ shielded PU sheath cable 3m	02	-----

Note- The quotation not fulfilling strictly the above technical specification will be rejected.

Satish

