

UNDER POSTING OF CERTIFICATE



Sardar Vallabhbhai National Institute of Technology,
Surat(GUJARAT)395007

Department of Chemical Engineering

Ref. No: DoChE/AAM/792/2026

Date: 07/08/2026

SUB:Enquiry for Supply of Peristaltic Pump (Digital).

You are invited to quote your prices for purchase of **Peristaltic Pump (Digital)**. The quotation may be sent in a sealed envelope with the superscription "**Quotation with reference to the Enquiry No. DoChE/AAM/ 792/2026**" addressed to **The Director, Sardar Vallabhbhai National Institute of Technology, Surat – 395007, Gujarat**, with kind attention to **Dr. Alka A. Mungray, Professor, Department of Chemical Engineering**. The quotation should reach the undersigned on or before **28/08/2026**.

The quotations should be furnish with the following information

1. The quotation for the above subject quote the rate and to be taken from our Institute includes labor charge, transportation etc. to take the time form various location of the institute.
2. If asked, samples are to accompany the quotations.
3. Tax, Custom duty, Insurance charges, packing and forwarding charges, if not included under the price quoted, should be clearly specified.
4. The period of validity of the quotation must be specified Offers subject to prior sale may please be avoided.
5. The supply period is to be clearly mentioned in the quotation.
6. The mode of delivery of the stores may be mentioned. The delivery should be F.O.R SVNIT, Surat.
7. All concessions available to an educational institution should be specified and also taken into account while quoting.
8. Payment will be made only after satisfactory of the supplied items.
9. Your specifications and terms conditions should be on your company letterhead & signed by an authorized person.
10. Offered Quotation may be rejected if any ambiguity is found in the offered specifications, terms & conditions supplied by the party in specified tabular format.
11. The Director/Institute reserves the right to accept stores which are not strictly in confirming with the specifications, but otherwise found suitable.

Yours faithfully,

Head, DoChE

Item Description

Sr. no.	Item	Quantity	Unit Price (Rs.)	Total Price (Rs.)
1	Peristaltic Pump, Digital Flow Range (0.005–641 mL/min)	02		

Detailed Specifications:

Peristaltic Pump (Digital)

- Flow range: **0.005–641 mL/min**
- Digital display with programmable flow rate and speed control.
- Flow accuracy: **±0.5 % or better.**
- Variable speed control with microprocessor-based drive.
- Suitable pump head for continuous laboratory operation.
- Compatible with standard laboratory-grade tubing
- Bidirectional operation (Clockwise and Counter-clockwise).
- Flow calibration facility.
- Continuous and timed dispensing modes.
- Complete unit with pump head, tubing, power cord and operating manual.
- Minimum **one-year comprehensive warranty.**

Handwritten signature
07/08/26